

Fire Engineering & Science Technology International Co., Ltd.

www.festi.co.kr

Control Point	
Rad O/P(mV)	Heat Flux (kW/m ²)
6.78	18.22
7.87	20.18
10.31	27.48
11.58	31.09

Gradient: 2.7837

Intercept: -1.1331

Residual: 0.9945

$$y = 2.7827x - 1.1331$$

Radiometer O/P at 25 kW/m² : 9.3928 mV

Client:

Calibration Engineer: Jeong-hyeon Choo

Calibration Date:

Copper Disc Calorimeter Serial No:

